

Japan Update

Yukina UENO

Deputy Director (International), Medical Device Evaluation Division, Pharmaceutical Safety Bureau,
Ministry of Health, Labour and Welfare Japan (MHLW)



Agenda

- Establishment of PMDA's International Hubs
- Software as a Medical Device
- Cybersecurity





Agenda

- **Establishment of PMDA's International Hubs**
- Software as a Medical Device
- Cybersecurity






PMDA's International Hubs



Asia Office,
Bangkok



PMDA Central
Office, Tokyo



Washington
D.C Office

***Establishment of PMDA's international hubs
to enhance international contribution/capability for regulatory proposal***

PMDA's 5th mid-term plan



Initiatives to strengthen cooperation with Asian countries / with the United States & EU

➤ To support **innovative medicines & medical devices access in Japan and Asian countries,**

- Strengthening cooperation with ASEAN countries
- Supporting promotion of regulatory harmonisation with Asian countries
- Developing an environment for smooth clinical development

➤ **Close collaboration between Japan, US and European regulatory authorities is essential in supporting;**

- Development/distribution of **innovative medicines and medical devices**
- Regulatory review
- Post-marketing measures



Establishment of
Asia Office
in Bangkok, Thailand



Establishment of
Washington D.C. Office,
USA

PMDA dispatches a staff to EMA as
a liaison officers (since 2009~).



Agenda

- Establishment of PMDA's International Hubs
- **Software as a Medical Device**
- Cybersecurity





Software as a Medical Device

DASH for SaMD 2 (2023/9/6)

- ◆ Organize and publicize the two-step approval scheme for SaMD
- ◆ **Develop guidelines for approval review and marketing procedures for SaMD for the general public**
- ◆ Promotion of overseas acceptance of our review results (such as English translation of review reports)
- ◆ Subsidies for development funds for SaMD developers
- ◆ Support for SaMD developers to actively business overseas

DASH for SaMD (2020/11/24)

- ◆ Setup an office to review SaMD in MHLW and PMDA
- ◆ Establishment of SaMD centralized consultation service
- ◆ Next-generation medical device evaluation index, development guidance, audit points, and certification criteria formulation
- ◆ Trial implementation of priority review, etc. for innovative SaMD
- ◆ Promote the use of IDATEN (Improvement Design within Approval for Timely Evaluation and Notice) and streamline procedures, etc.

<Expand and continue>

- ◆ Upgrade from office to Department for reviewing SaMD in PMDA
- ◆ Establishment of SaMD-specific consultation service
- ◆ (Continue)
- ◆ (Continue)
- ◆ (Continue)

Review point for;

- Software for Peritoneal Dialysis Treatment
- Supporting Software for Dental Implant Treatment
- Software for Ophthalmic Surgery Treatment Planning
- Supporting Software for Detecting Lesion with Endoscopic Imaging
- Computer-Aided Diagnosis Program to Support Interpretation of Medical Images





Collaborative forum among regulator, academia and industry (SaMD Forum)

Background of SaMD Forum

- The Ministry of Health, Labor and Welfare (MHLW) and the Ministry of Economy, Trade and Industry (METI) established the “SaMD Industry-Academia-Government Forum” from FY2021 to exchange opinions among industry, academia and government on the identification of the seed technologies of SaMD and on issues and solutions concerning and organization of the development method.



Collaborative forum among regulator, academia and industry (SaMD Forum)

Date	Forum	Theme of discussion
February 4, 2022	The 1st SaMD Forum	Collaboration among regulator, academia and industry
December 1, 2022	The 2nd SaMD Forum	Issues in predictability of business monetization and utilization of medical data to promote development of SaMDs
February 6, 2023	SaMD Sub-forum 2023	Ideal medical remuneration system for utilization of SaMDs and utilization of medical data in SaMDs
September 21, 2023	The 3rd SaMD Forum	Ideal regulatory approval system and insurance system according to the characteristics of SaMDs
February 7, 2024	SaMD Sub-forum 2024	Issues in overseas expansion of SaMDs and AMED's support measures for development of SaMDs
September 3, 2024	The 4th SaMD Forum	Issues in overseas expansion of SaMDs and domestic implementation of the value of SaMDs
February 10, 2025	SaMD Sub-forum 2025	Regulation of AI-based SaMDs and progress report of "DASH for SaMD2"



Agenda

- Establishment of PMDA's International Hubs
- Software as a Medical Device
- **Cybersecurity**





Cybersecurity

<Article 1-11, p.,1, Enforcement Regulations of the Medical Care Act>
 Pursuant to Article 6-12 of the Law, administrators of hospitals and other facilities must **ensure the following systems for safety management:**
 - Establish guidelines for safety management related to medical care.

<Article 14, p.,1, Enforcement Regulations of the Medical Care Act>
 The administrator of a hospital or clinic must **take necessary precautions** to ensure that the drugs, medical devices, and regenerative medicine products present in the hospital or clinic do **not violate the provisions of the Pharmaceuticals and Medical Devices Act.**


<Article 14, p.,2, Enforcement Regulations of the Medical Care Act>
 The administrator of a hospital, clinic, or midwifery clinic must take the necessary measures to **ensure cybersecurity (meaning cybersecurity as defined in Article 2 of the Basic Act on Cybersecurity)** so as to avoid any significant disruption to the provision of medical care.

<Guidelines for the security management of medical information systems*>
 Medical institutions will take the initiative **in managing the security of medical information systems to ensure their confidentiality, integrity, and availability.**

< Essential Principles**, Article12-3 >
 Medical device manufacturers to play a key role in **maintaining the functionality of medical devices and patient safety against cyber risks.**
 Provide necessary information to medical institutions and cooperate with them.


*Guidelines for the Security Management of Medical Information Systems Version 6.0 (May 2023)

** Standards for medical devices established by the Minister of Health, Labor and Welfare pursuant to the provisions of Article 41, Paragraph 3 of the Act on Ensuring the Quality, Efficacy, and Safety of Pharmaceuticals, Medical Devices, etc. (Ministerial Notification No. 122 of 2005, Partial amendment of Ministerial Notification No. 403 of 2014, Partial amendment of Ministerial Notification No. 67 of 2020)






Cybersecurity



English information available



Thank you for your attention